

Anthraquinone dye

Experimental

Column: SVEA Core C18, 2.6 μ m, 4.6*100mm

Instrument: HPLC


Flow rate: 1.0 mL/min

Column temperature: 40 °C

Detector: UV 250 nm

Sample:
1 Alizalin
2 Chrysazin
3 Anthrarufin

Mobile phase: Methanol/20mM phosphate buffer pH 2.(75/25)



Mobile phase: Methanol/0.1% formic acid (75/25)

