

Maleic and fumaric acids

Experimental

Column: SVEA CN, 5 μm 110 \AA 4.6 x 150 mm
Instrument: HPLC
Mobile phase: MeOH/20 mM ammonium acetate buffer pH 3.9 (03/97)
Flow rate: 1.0 mL/min
Column temperature: 40°C
Detector: UV 210 nm
Sample:
1. Fumaric acid
2. Maleic acid

