

Tropisetron

Experimental

Column: SVEA C18 Gold, 5 µm 110 Å 4.6 x 150 mm

Instrument: HPLC


Flow rate: 2.0 mL/min

Column temperature: 30 °C


Detector: UV 254 nm

Sample:
1. Des-Me-Tropisetron
2. Tropisetron

Mobile phase: 0.1% TFA in water pH 1.9/methanol (65/35)




1



Tropisetron

2



des-Me-Tropisetron

Mobile phase: A: 0.1% TFA in water pH 1.9; B: Methanol.
Gradient: 5-95% B in 18 minutes

