

Dimethyl fumarate

Experimental

Column:	SVEA Gold C18, 5 µm 110 Å 4.6 x 150 mm
Instrument:	HPLC
Mobile phase:	Methanol: water = 40:60 (v: v)
Flow rate:	1.0 ml/min
Injection volume:	10 µl
Sample concentration:	5 µg/ml
Column temperature:	35 °C
Detector:	UV 220 nm
Sample:	Dimethyl fumarate

