

Sodium formaldehyde sulfoxylate

Experimental

Column:	SVEA Gold C18, 5 μ m 110 Å 4.6 x 150 mm
Instrument:	HPLC
Mobile phase:	Methanol: water = 65:35 (v: v)
Flow rate:	1.0 ml/min
Injection volume:	10 μ l
Sample concentration:	10 μ g/ml
Column temperature:	35 °C
Detector:	UV 478 nm
Sample:	Formaldehyde

