

Vitamin K1

Experimental

Column:	SVEA Gold C18, 5 μm 110 \AA 4.6 x 150 mm
Instrument:	HPLC
Mobile phase:	Methanol: water = 98:2
Flow rate:	1.0 ml/min
Injection volume:	10 μl
Sample concentration:	20 $\mu\text{g/ml}$
Column temperature:	50 $^{\circ}\text{C}$
Detector:	UV 248nm
Sample:	Vitamin K1

